

Mapping for Digital 128-Channel Probe

SPI connector

ASSY-350 H20

Electrodes are located on the side of the probe as indicated by the arrow below - use Shank A orientation as a reference to construct the channel-map with respect to your brain target. Electrode numbers are 0-indexed so as to correspond to the Intan / Open Ephys data-acquisition software channels.

Red Wire and White Wire correspond to indicated REF and GND inputs to the headstage amplifier.

Please note that channels highlighted in red are offset vertically according to the schematic on the left

